

MASS SPECTROMETRY REPORT

Selank

Lot 20260313

MANUFACTURE DATE

March 13, 2026

ANALYSIS DATE

March 14, 2026

REPORTED PURITY

99.45%

DOCUMENT TYPE

Mass Spectrometry Report

This document is provided as part of Kodiak's lab transparency archive. It accompanies a research-use-only compound and reflects vendor analytical documentation. Independent third-party verification by Janoshik Analytical is performed on rotating samples and archived alongside vendor reports.

The following pages reproduce the original report as received from the manufacturer or analytical laboratory. No content has been altered.


ARCHIVE REFERENCE

kodiak-ms-selank-20260313.pdf

PUBLISHED BY

kodiak.org

MASS SPECTROMETRY REPORT



Sample Description

Analyzed date: 2026-03-13
Analyst: PY
Sample: DZT-043
M.W.: 751.87
Lot. No.: 20260313-01

Instrument

Probe:
Nebulizer Gas Flow:
CDL:
CDL Temp.:
Block Temp.:

Waters ZQ2000

ESI
1.5L/min
-20.0v
250 °C
200 °C
Probe Bias: +4.5kv
Detector: 1.5kv
T. Flow: 0.2ml/min
B. Conc.: 50%H2O/50%ACN